TEST REPORT

(Electronic version)

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: KCBU-4454-54

No:200048889

ISSUE DATE:2020-03-27

APPLICANT: NEW INDUSTRIES HEALTH SCIENCE AND TECHNOLOGY (ZHUHAI) CO.,LTD
ADDRESS: SECOND FLOOR OF 8017 ZHUHAI AVENUE, PINGSHA TOWN, ZHUHAI CITY, GUANGDONG PROVINCE

APPLICANT PROVIDED SAMPLE DESCRIPTION:

FIFTY(50) PIECES OF DISPOSABLE PROTECTIVE MASK

STYLE NO.: 200302

SIZE: 10PIECE/BAG 17.5cm×9.5cm

DATE RECEIVED/DATE TEST STARTED: 2020-03-24

CONCLUSION:

APPEARANCE QUALITY[3 PIECES]	M
NOSE CLIP[3 PIECES]	M
MARKING	M
SIZE DEVIATION[3 PIECES]	M

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:
THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200048031.
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
THE EXPERIMENT WAS CARRIED OUT AT No.1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P.R.CHINA.

APPROVED BY: 刘圆

Yuan Liu ENGINEER

总部： 广州市番禺区珠江路1号
花都实验室： 广州市花都区狮岭镇旗岭河滨西路1号
电话:020-61994598/61994599
电话:020-37721161

TEST REPORT

(Electronic version)

No:200048889



总部： 广州市番禺区珠江路1号
花都实验室： 广州市花都区狮岭镇旗岭河滨西路1号

电话:020-61994598/61994599
电话:020-37721161

TEST REPORT

(Electronic version)

No:200048889

APPEARANCE QUALITY[3 PIECES]

(Q/ZHXCYDJK001-2020 6.2)

PASS	REQUIREMENT THE APPEARANCE OF THE MASK SHOULD BE NEAT AND INTACT, AND THERE SHOULD BE NO DAMAGE OR STAIN ON THE SURFACE. (Q/ZHXCYDJK001-2020)

NOSE CLIP[3 PIECES]

(Q/ZHXCYDJK001-2020 6.4.1)

NOSE CLIP		REQUIREMENT THE PLASTIC NOSE CLIP SHOULD BE ATTACHED TO THE MASK. THE NOSE CLIP LENGTH SHOULD NOT BE LESS THAN 8.0cm. (Q/ZHXCYDJK001-2020)
PASS		
NOSE CLIP LENGTH		
1#	10.5cm	
2#	10.4cm	
3#	10.4cm	

MARKING

(Q/ZHXCYDJK001-2020 8)

PASS	REQUIREMENT ACCORDING TO THE CLAUSE 8.1 OF THE PRODUCT STANDARD (Q/ZHXCYDJK001-2020)

REMARK: THE ITEM DOES NOT INCLUDE VERIFICATION OF THE AUTHENTICITY OF THE CONTENT.

SIZE DEVIATION[3 PIECES]

(Q/ZHXCYDJK001-2020 6.3)

PASS		REQUIREMENT THE MASK SHOULD CONFORM TO THE DESIGN SIZE, THE MAXIMUM TOLERANCE SHOULD NOT EXCEED ±5%. (Q/ZHXCYDJK001-2020)
MAXIMUM SIZE DEVIATION	-2%	

———End of Report———

总部： 广州市番禺区珠江路1号
花都实验室： 广州市花都区狮岭镇旗岭河滨西路1号
电话:020-61994598/61994599
电话:020-37721161